

## 資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成28年10月21日

契約担当役

独立行政法人労働者健康安全機構

福島労災病院 院長 渡辺 毅

◎調達機関番号 590 ◎所在地番号 07

### ○第84号

#### 1 調達内容

- (1) 品目分類番号 71、27
- (2) 導入計画物品及び数量  
医用画像情報システム 一式
- (3) 調達方法 借入
- (4) 導入予定時期 平成29年度第2・四半期以降
- (5) 調達に必要なとされる基本的な要求要件
  - ① 本システムは、独立行政法人労働者健康安全機構福島労災病院における診療業務及び経営管理業務を総合的に支援するためのもので、画像撮影装置から受信した画像データを統合的に保存、閲覧、管理、配信できること。
  - ② 本システムは、DICOM規格に準拠していること。
  - ③ 本システムは、当院の保有する病院情報システム、放射線部門業務システム等と連携可能であり、患者基本情報や診療情報、画像データなどを送受信できること。
  - ④ 本システムは、当院の委託する遠隔画像診断支援サービス業者と連携できること。
  - ⑤ 本システムは、既存の診療用画像データ、レポートデータなどを継承し、継続して閲覧、使用できること。
  - ⑥ 本システムは、病院経営、診療業務を高度に支援するため、最新の技術に基づくこと。また、操作性及び応答時間は、十分実業務に対応できること。
  - ⑦ 本システムのうち、病院業務に不可欠なシステムは、24時間、無人であっても自動稼働し、安全で容易に運用できること。また、システム障害対策及び情報セキュリティ対策が十分になされていること。
  - ⑧ 本システムは、今後病院情報システムのベン

データ変更等があった場合でも、継続的に使用することができる拡張性の高い仕様であること。

- 2 資料及びコメントの提供方法 上記1 (2) の物品に関する一般的な参考資料及び同 (5) の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
  - (1) 資料等の提供期限 平成28年12月5日15時00分（郵送の場合は、前日迄に必着のこと。）
  - (2) 提出先  
〒973-8403 福島県いわき市内郷綴町沼尻3番地  
福島労災病院会計課契約係 加藤 健雄  
電話0246-26-1111
- 3 説明書の交付 本公表に基づき応募する供給者に対して導入説明書を交付する。
  - (1) 交付期間 平成28年10月21日から平成28年12月5日まで。
  - (2) 交付場所 上記2 (2) に同じ
- 4 その他 この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
- 5 Summary
  - (1) Classification of the services to be procured: 71, 27
  - (2) Nature and quantity of the services to be required: Picture Archiving and Communication System 1 set.
  - (3) Type of the procurement: Lease
  - (4) Basic requirements of the procurement:
    - ① This system should generally support the medical care services and business management at Japan Organization of Occupational Health and Safety Fukushima Rosai Hospital. This should be an integrated system which can save, browse, control and deliver the image information which the medical imaging apparatus send.
    - ② This system should conform to the DICOM standard.
    - ③ This system should cooperation with the present network system of the hospital information system and radiology information system, and will send and receive the patient basic information, medical care records and image information.
    - ④ This system should work with the remote

diagnostic image reading and reporting service provider who Fukushima Rosai Hospital entrusts.

- ⑤ This system should inherit the existing consultation image information and reporting data, and will continue to browse and use.
  - ⑥ This system must be based on the latest technology to support hospital business and medical services. And the operability and response time must be appropriate for the actual medical care.
  - ⑦ The part of this system which is indispensable for the hospital business must be operated automatically for 24 hours a day, even if it is unmanned, and always operated safely and easily, Also it should take enough measures to cope with the system failure and information security.
  - ⑧ This system should have highly scalable specification that is possible to use continuously even if the vendor of hospital information system change in future.
- (5) Time-limit for the submission of the requested material: 3:00 PM, December 2016
- (6) Contact point for the notice: Takeo Katou Procurement Section, Accounting Division, Fukushima Rosai Hospital 3 Numajiri Tudura-machi Uchigo Iwakishi Fukushima-ken 973-8403 Japan TEL 0246-26-1111.